

(19)



Europäisches Patentamt

European Patent Office

Office européen des brevets.



(11)

EP 0 943 288 A1

(12)

EUROPEAN PATENT APPLICATION

(43) Date of publication:
22.09.1999 Bulletin 1999/38

(51) Int Cl.⁶: A61B 3/10, G01B 11/24

(21) Application number: 99302096.5

(22) Date of filing: 18.03.1999

(84) Designated Contracting States:
AT BE CH CY DE DK ES FI FR GB GR IE IT LI LU
MC NL PT SE
Designated Extension States:
AL LT LV MK RO SI

(72) Inventors:
• Suzuki, Takayoshi
Hamamatsu-shi, Sizuoka-ken 431-2103 (JP)
• Yokoi, Norihiko, 251-1 Kamigoryo Kamiecho
Kyoto-shi, Kyoto-Fu 603-8147 (JP)

(30) Priority: 20.03.1998 JP 7161598

(74) Representative: Sturt, Clifford Mark et al
Miller Sturt Kenyon
9 John Street
London WC1N 2ES (GB)


(71) Applicants:
• KOWA COMPANY LTD.
Nagoya-shi Aichi-ken, 460-8625 (JP)
• Yokoi, Norihiko
Kyoto-shi, Kyoto-Fu 603-8147 (JP)

(54) Ophthalmic apparatus

(57) A grid (27) formed with apertures of a predetermined shape is illuminated with light from a lamp (21) and projected onto a tear meniscus (28) accumulated on the lower eyelid (36). The aperture image projected on the tear film is imaged by a CCD camera (31). The tear meniscus functions as a concave mirror, so that the

magnification factor depends on the radius of meniscus curvature. A processor (32) calculates the magnitude of the grid image and evaluates the radius of meniscus curvature. Since the volume of lacrimal fluid (V_1, V_2) varies depending upon the radius of meniscus curvature (r_1, r_2), the latter is used as a value representing the lacrimal fluid volume for dry eye diagnose purposes.

FIG. 3



BEST AVAILABLE COPY

Description

[0001] The present invention relates to an ophthalmic apparatus, and more particularly to an ophthalmic apparatus for non-contact measurement of the physical quantity of lacrimal fluid collected on the lower eyelid.

[0002] Recent years have seen an increase in the numbers of people suffering from dry eyes caused by working with VDTs (visual display terminals) or working in rooms in which the air is dried out by air conditioning systems. Dry eyes can result in a number of ophthalmic conditions, such as damage to the corneal epithelium and conjunctiva. As such, the diagnosis of dry eye syndrome is becoming an important part of ophthalmic diagnostic procedure.

[0003] Conventional methods of diagnosing dry eye include examining vital stainings and volume of lacrimal fluid. However, such methods involve discomfort to the patient caused by the application of a solution or contacting the eye with an instrument. To detect the dry eyes in a non-contact manner, methods have been tried involving projecting a beam of coherent light onto the eye and examining interference fringes formed by the tear film layer. In the apparatus of such systems, color images of interference fringes (rainbow-colored interference patterns) formed by the tear film lipid layer of an eye to be examined are photoelectrically converted by a photoelectric element in a light-receiving system and shown on a display means. The presence of dry eye can then be readily diagnosed by examining the interference pattern indicating the condition of the tear film layer.

[0004] However, a problem with the interference fringes produced by tear film lipid layer with the conventional systems is the low contrast of the fringes, which makes it difficult to obtain a good ophthalmic diagnosis based on the fringes. Another problem is that the examiner directly observes the color patterns on the display screen to evaluate the grade of the dry eye condition, so only qualitative measurement is possible.

[0005] The object of the present invention is to provide an ophthalmic apparatus that enables diagnosis of dry eye condition by quantitatively measuring the physical quantity of lacrimal fluid collected on the border of the lower eyelid.

[0006] In accordance with the present invention, the above object is attained by an ophthalmic apparatus comprising a light control means formed with an aperture having a predetermined shape, means for projecting the aperture onto a surface of tear film collected on a lower eyelid, means for imaging the aperture projected on the tear film surface, and a means for evaluating a physical quantity of lacrimal fluid based on the image of the aperture thus obtained.

[0007] The tear film surface on the border of the lower eyelid (tear meniscus) functions like a concave mirror, so the magnification factor of the imaged aperture depends on the radius of meniscus curvature. In this invention, the magnification factor of the aperture image

is obtained, giving the radius of meniscus curvature. The radius of meniscus curvature has a bearing on the volume of lacrimal fluid affecting the dry eye condition, so obtaining the radius of meniscus curvature makes it possible to evaluate quantitatively the degree of severity of the dry eye condition, or the phase into improvement thereof.

[0008] The above and other features of the present invention will become apparent from the following description made with reference to the drawings.

Figures 1a and 1b are diagrams illustrating the state of tear film collected in the lower eyelid, and Figures 1c and 1d are diagrams illustrating the relationship between the radius of meniscus curvature and the volume of lacrimal fluid;

Figure 2a is an optical diagram of the principle of the measurement of the radius of meniscus curvature, Figure 2b is a diagram illustrating the formation of a grid image by the tear meniscus, and Figure 2c is a diagram of the grid image displayed on a monitor;

Figure 3 is a diagram of the arrangement of the apparatus of the invention;

Figure 4 is a diagram of an example of grid dimensions;

Figure 5 is a diagram of another example of grid dimensions; and

Figure 6 is a diagram of the grid image projected onto the center and peripheral portions of the tear meniscus.

[0009] Figure 1 illustrates the principle of a meniscometer for measuring the quantity of lacrimal fluid. A tear film from the outermost layer of the eye is comprised of lipid, aqueous and mucin layers; the lipid layer is secreted from the Meibomian gland and is subjected to pressure when the eyelid closes. When this happens, a tear film 3 accumulates on the lower eyelid border, as shown in Figures 1a and 1b. An examination for dry eye is conducted with respect to the volume of lacrimal fluid on the border of the lower eyelid, with less lacrimal fluid indicating a more severe dry eye condition.

[0010] The relationship between tear volume V and radius of meniscus curvature r of the surface of the tear film 3 is that, since the cornea 4 and the eyelid 2 are both curved, a larger V results in a larger r . This relationship is illustrated by Figures 1c and 1d. If lacrimal fluid volume V increases from V_1 to V_2 , the radius of meniscus curvature also increases from r_1 to r_2 . Here, θ_1 , θ_2 are constants determined by the surface tension. While these constants vary according to the severity of the dry eye condition, compared to the change in r , such change is negligibly small.

[0011] In the present invention, the volume V of lacrimal fluid is determined by measuring the radius of meniscus curvature r (tear meniscus). For this, in accordance with the invention, a grid image is projected onto

the tear film surface, and the physical quantity of the lacrimal fluid, that is, the radius of meniscus curvature r , is measured by analyzing the grid image. Figure 2 shows the configuration of this principle.

[0012] In Figure 2a, light from a lamp 10 passes through an objective lens 11 and illuminates a grid 12, used as a light control means, which is projected onto a concave mirror 13 serving as a tear meniscus model. A grid image 14 is formed by the concave mirror 13, and, via a projection lens 15, this grid image 14 is imaged by an imaging means such as a camera 16.

[0013] With respect to Figure 2b, the grid having a height d at a working distance W from the concave mirror (tear meniscus) of curvature radius r is transformed into an image d_1 at a distance W_1 from the concave mirror. Here, $W_1 = (rW)/(2W - r)$ is obtained from the well-known relationship $2/r = (1/W) + (1/W_1)$, and $d_1 = (d/W) \times W_1$ is obtained from the relationship $d_1/d = W_1/W$.

[0014] From the above two equations,

$$d_1 = (d/W) \times \{rW/(2W - r)\},$$

$$d_1 = \{dr/(2W - r)\}.$$

[0015] Here, if $W \gg r$ (for example, $W = 24$, $r = 0.3$), then $2W - r \approx 2W$, providing the approximation formula $d_1 \approx (dr/2W)$, thus $r = (d_1/d) \times 2W$.

[0016] If grid image d_1 is magnified by a magnification of β to the size D shown in Figure 2c, since $D = \beta \times d_1$, then $r = (D/\beta) (2W/d)$. Thus, if the monitor is a 14-inch television monitor, for example, then $\beta = 190.9$, and as a result, $r = (D/190.9) (2W/d)$.

[0017] In grid size (grid pitch), d is a constant, and W is the working distance value determined by the design. While this might be changed somewhat in the alignment, it is a small enough value to be disregarded. Thus, the curvature radius r of the tear meniscus for lacrimal fluid volume V can be found by measuring the size of the grid image D displayed on the monitor.

[0018] Figure 3 illustrates a specific apparatus based on the principle described above. With reference to Figure 3, light from a halogen lamp 21 passes through a filter 22, which blocks heat rays, a polarizing plate 23 and an illumination lens 24, and is reflected by a mirror 25 such as a half-mirror or aperture mirror. The light reflected by the mirror 25 passes through an objective lens 26 and illuminates a grid 27 having a plurality of apertures, that functions as a light control means. The grid 27 thus illuminated is projected onto a meniscus 28 of tear film collected on a lower eyelid 36.

[0019] As shown in Figure 4, the grid 27 comprises a plurality of apertures in the form of slits (five in this example) each measuring D_3 by D_5 arranged on a D_1 by D_2 substrate, with the slits being equidistantly spaced apart by a distance D_4 , and provided with a white space D_6 above and below and a white space D_7 on each side.

The dimensions are set at, for example, the following: $D_1 = 48.0$ mm, $D_2 = 15.0$ mm, $D_3 = 9.0$ mm, $D_4 = 4.0$ mm, $D_5 = 4.0$ mm, $D_6 = 6.0$ mm, and $D_7 = 3.0$ mm. In this example, the d of Figure 2b (grid pitch) is $d = D_4 + D_5 = 8$ mm.

[0020] The light from the projected grid 27 is reflected by the tear meniscus 28, forming a grid image in the vicinity of the tear meniscus. The grid image thus formed by the tear meniscus passes through objective lens 26, mirror 25, projection lens 29 and polarizing plate 30 and is picked up by a CCD camera 31, and the image is subjected to image processing by a processor 32. This processor 32 can, for example, be used to obtain the pitch of the grid image on the camera corresponding to $d = D_4 + D_5$ by binarizing the image signal and obtaining the pixel coordinates for each aperture. The processor 32 also calculates the pitch d_1 of the grid image formed by the tear meniscus, taking into account the lens magnification factor, and evaluates the radius of curvature r of the tear meniscus 28 in accordance with the above equation $r = (d_1/d) \times 2W$.

[0021] As described above, in the formation of the grid image, the tear meniscus 28 has the function of a concave mirror, and, therefore, the factor by which the grid image formed is magnified depends on the radius of meniscus curvature r of the tear meniscus. Obtaining the radius of meniscus curvature makes it possible to evaluate quantitatively the severity of the dry eye condition. The outcome of each calculation and the evaluation can be displayed on a monitor 33.

[0022] The polarizing plate 23 arranged in the illumination and projection system and the polarizing plate 30 arranged in the imaging system both have the same orientation so as to transmit light in the same direction. As the tear meniscus is liquid, the polarized state is not readily broken down in the course of reflection, so using the polarizing plates makes it possible to improve the signal-to-noise (S/N) ratio during imaging.

[0023] Figure 5 shows another example of a grid. In this example, the grid has a finer pitch. The dimensions are $D_1 = 48.0$ mm, $D_2 = 15.0$ mm, $D_3 = 9.0$ mm, $D_4 = 2.0$ mm, $D_5 = 2.0$ mm, $D_6 = 7.0$ mm, and $D_7 = 3.0$ mm, and the grid pitch will be 4 mm.

[0024] As to what the degree of precision of r is when the working distance W is 24 mm, the grid pitch is the 8 mm of Figure 4, and the monitor 33 is a 14-inch model, the following is the result of an actual measurement carried out with the apparatus of Figure 3 (not using the processor 32), using a glass tube of radius 0.30 mm and piano wire of radius 0.15 mm. In the case of a 14-inch monitor, $\beta = 190.9$, so from the above equation, curvature radius r will be as follows.

$$r = (D/190.9) \times \{(2 \times 24)/8\} = 0.0314 \times D.$$

[0025] In the case of the glass tube of 0.30 mm radius, the grid pitch D on the monitor was 9.55 mm, this being

the mean value of ten measurements obtained using a ruler, so $r = 0.0314 \times 9.55 = 0.30$ mm, an accurate value. In the case of the piano wire of radius 0.15 mm, the average of ten measurements of D was 4.58 mm, so $r = 0.0314 \times 4.58 = 0.14$ mm. The degree of error is a mere 0.01 mm, confirming that the curvature radius of the meniscus can be measured with quite a degree of precision.

[0026] The tear meniscus 28 is a horizontally elongated shape, with the angle changing going toward the outside corner of the eye. So, by making the grid 27 rotatable, as shown in Figure 6, the orientation of the grid 27 can be changed between the middle grid image 40 projected on the tear film surface and the peripheral grid image 41. For example, the grid could be adjusted so that straight lines at right-angles to the grid apertures come to a point, making it possible to efficiently direct the illuminating light onto the tear meniscus.

[0027] As described in the foregoing, in accordance with the present invention, physical quantities such as the radius of meniscus curvature of tear film can be calculated based on a grid aperture image projected onto the tear film surface, thereby making it possible to quantitatively evaluate the degree of severity or change of a dry eye condition.

Claims

1. An ophthalmic apparatus comprising:
 - a light control means (27) formed with an aperture having a predetermined shape;
 - means (24,25,26) for projecting the aperture onto a surface of tear film (28) collected on a lower eyelid (36);
 - means (25,26,29) for imaging the aperture projected on the tear film surface; and
 - a means (32) for evaluating a physical quantity of lacrimal fluid (V_1, V_2, r_1, r_2) based on the image of the aperture thus obtained.
2. An ophthalmic apparatus according to claim 1, wherein a radius of meniscus curvature (r_1, r_2) is calculated based on the aperture image.
3. An ophthalmic apparatus according to claim 2, wherein a dry eye condition is evaluated based on the calculated radius of meniscus curvature (r_1, r_2).
4. An ophthalmic apparatus according to any of claims 1 to 3, wherein the light control means is a grid (27) comprised of a plurality of slit-shaped apertures arranged in an equidistant arrangement.
5. An ophthalmic apparatus according to any of claims 1 to 4, wherein the projection means and imaging means are each provided with a polarizing plate
6. An ophthalmic apparatus according to any of claims 1 to 5, wherein an orientation of an aperture (40,41) can be adjusted in accordance with a position of the aperture projected on the tear film surface.
7. An ophthalmic apparatus according to any of claims 1 to 6, wherein an optical system of the projection means (25,26) and an optical system of the imaging means (25,26) are arranged coaxially.

BEST AVAILABLE COPY

FIG. 1 a




FIG. 1 b




FIG. 1 c





FIG. 1 d



BEST AVAILABLE COPY







FIG. 2c



BEST AVAILABLE COPY

FIG. 3






FIG. 4

FIG. 6





FIG. 5

BEST AVAILABLE COPY



European Patent
Office

EUROPEAN SEARCH REPORT

Application Number
EP 99 30 2096

DOCUMENTS CONSIDERED TO BE RELEVANT			
Category	Citation of document with indication, where appropriate, of relevant passages	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int.Cl.6)
A	GB 2 123 977 A (SMITH AND NEPHEW ASSOCIATED COMPANIES) 8 February 1984 * page 1, line 5 - line 13 * * page 1, line 44 - line 46 * * page 2, line 61 - page 3, line 23 *	1-4,7	A61B3/10 G01B11/24
A	JULIA C. MAINSTONE ET AL.: "Tear meniscus measurement in the diagnosis of dry eye" CURRENT EYE RESEARCH, vol. 15, no. 6, June 1996, pages 653-661, XP002105229 * abstract * Sections "Apparatus and procedures" and "Tear meniscus image analysis" * page 655 - page 656 *	1-3	
A	US 5 719 659 A (SUZUKI TAKAYOSHI) 17 February 1998 * column 2, line 54 - column 3, line 18 *	1,5,7	
A	US 4 582 404 A (C. B. HAMILTON) 15 April 1986 * column 3, line 19 - line 65 * * column 5, line 18 - line 27 * * column 6, line 16 - line 26 *	1,2,6	TECHNICAL FIELDS SEARCHED (Int.Cl.6) A61B G01B
The present search report has been drawn up for all claims			
Place of search THE HAGUE		Date of completion of the search 8 June 1999	Examiner Knüpling, M
CATEGORY OF CITED DOCUMENTS X : particularly relevant if taken alone Y : particularly relevant if combined with another document of the same category A : technological background O : non-written disclosure P : intermediate document T : theory or principle underlying the invention E : earlier patent document, but published on, or after the filing date D : document cited in the application L : document cited for other reasons & : member of the same patent family, corresponding document			

EPO FORM 1503 03/92 (P04C01)

BEST AVAILABLE COPY

**ANNEX TO THE EUROPEAN SEARCH REPORT
ON EUROPEAN PATENT APPLICATION NO.**

EP 99 30 2096

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report.
The members are as contained in the European Patent Office EDP file on
The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

08-06-1999

Patent document cited in search report		Publication date	Patent family member(s)	Publication date
GB 2123977	A	08-02-1984	NONE	
US 5719659	A	17-02-1998	JP 10033483 A	10-02-1998
US 4582404	A	15-04-1986	NONE	

EPO FORM P0439

For more details about this annex : see Official Journal of the European Patent Office, No. 12/82

BEST AVAILABLE COPY